

Artikel: AB 6300-06
 Los: 8900-06
 Stück: 5
 Kunde: Textil AG

Datum/Zeit: 30.06.2006, 16:04 - 30.06.2006, 16:09
 Bed. Person: J. Schmidt


Dynamische Luftdurchlässigkeit

Startvolumen: 404 cm³
 Startdruck: 400 kPa
 Grenzdrücke: 30/70 kPa
 Prüffläche: 50 cm²


Datum/Zeit: 30.06.2006, 16:04 - 30.06.2006, 16:09
 Instrument: Textest FX 3350, S/N: 102

	MDLD	Exponent	
1:	2'010 mm/s	0.788	
2:	2'390 mm/s +	0.758	
3:	2'010 mm/s	0.781	
4:	3'020 mm/s +	0.788	««« gelöscht Schadstelle
4:	2'190 mm/s	0.798	

Kommentar:



MW:	2'150 mm/s	Soll:	2'000 mm/s
Min:	2'010 mm/s	Min:	1'700 mm/s
Max:	2'390 mm/s +	Max:	2'300 mm/s
CV:	8.4 %	Anzahl:	4
VB:	13.4 %	VB:	



MW:	0.781	Soll:	0.775
Min:	0.758	Min:	0.750
Max:	0.798	Max:	0.800
CV:	2.2 %	Anzahl:	4
VB:	3.5 %	VB:	

